




Tapentadol

The Drug Enforcement Administration's Special Testing and Research Laboratory generated this monograph using structurally confirmed reference material.



1. GENERAL INFORMATION

IUPAC Name: *rel*-3-[(2*R*,3*R*)-1-(dimethylamino)-2-methylpentan-3-yl]phenol

CAS#: 175591-09-0 (HCl)

Synonyms: Nucynta, Palexia, Tapal

Source: DEA Reference Material Collection

Appearance: White powder

UV_{max}(nm): Not determined

2. CHEMICAL AND PHYSICAL DATA

2.1 CHEMICAL DATA

Form	Chemical Formula	Molecular Weight	Melting Point (°C)
Base	C ₁₄ H ₂₃ NO	221	Not Determined
HCl	C ₁₄ H ₂₃ NO HCl	257	Not Determined



Tapentadol

The Drug Enforcement Administration's Special Testing and Research Laboratory generated this monograph using structurally confirmed reference material.




3. QUALITATIVE DATA

3.1 NUCLEAR MAGNETIC RESONANCE

Sample Preparation: Dilute analyte to ~10 mg/mL in DMSO- d_6 containing TMS for 0 ppm reference and dimethylsulfone as quantitative internal standard.

Instrument: 400 MHz NMR spectrometer
Parameters: Spectral width: at least containing -3 ppm through 13 ppm
Pulse angle: 90°
Delay between pulses: 45 seconds

^1H NMR: Tapentadol HCl; Lot 1665.1B1.1; DMSO; 400MHz





Tapentadol

The Drug Enforcement Administration's Special Testing and Research Laboratory generated this monograph using structurally confirmed reference material.



3.2 GAS CHROMATOGRAPHY/MASS SPECTROMETRY


Sample Preparation: Dilute analyte ~4.8 mg/mL in 1:1 CHCl₃: MeOH

Instrument: Agilent gas chromatograph operated in split mode with MS detector
Column: DB-5 MS (or equivalent); 15m x 0.25 mm x 0.25 μm
Carrier Gas: Helium at 1.5 mL/min
Temperatures: Injector: 280°C
MSD transfer line: 280°C
MS Source: 250°C
MS Quad: 150°C
Oven program:
1) 100°C initial temperature for 1.0 min
2) Ramp to 280°C at 12 °C/min
3) Hold final temperature for 9.0 min

Injection Parameters: Split Ratio = 25:1, 1 μL injected
MS Parameters: Mass scan range: 30-550 amu
Threshold: 150
Tune file: stune.u
Acquisition mode: scan

Retention Time: 6.73 min

EI Mass Spectrum: Tapentadol HCl Lot 1665.1B1.1





Tapentadol

The Drug Enforcement Administration's Special Testing and Research Laboratory generated this monograph using structurally confirmed reference material.





3.3 INFRARED SPECTROSCOPY (FTIR)

Instrument: FTIR with diamond ATR attachment (1 bounce)

Scan Parameters:
Number of scans: 32
Number of background scans: 32
Resolution: 4 cm^{-1}
Sample gain: 1
Aperture: 150

FTIR ATR (Diamond 1 Bounce): Tapentadol HCl Lot 1665.1B1.1





Tapentadol

The Drug Enforcement Administration's Special Testing and Research Laboratory generated this monograph using structurally confirmed reference material.



4. ADDITIONAL RESOURCES

[*Wikipedia*](#)

[*Forendex*](#)